


## User's Manual

# Wing Aluminium trestle with adjustable heights



Guarantee of Quality system for the production and the final control of the products certified by the notifying body  
TÜV Italia S.r.l.

### INDEX

<b>General Information</b>	<b>page 2</b>	<b>Operating Instructions</b>	<b>page 3</b>
<b>Warnings</b>	<b>page 2</b>	<b>Maintenance and Cleaning</b>	<b>page 4</b>
<b>Description of Product</b>	<b>page 3</b>	<b>Accessories and Spare parts</b>	<b>page 4</b>

Spencer Italia S.r.l. Strada Cavi, 7 - 43044 Collecchio (Parma) - ITALY  
tel. 0039 0521 541111 - fax 0039 0521 541222 - e-mail: info@spencer.it

# Thank you for choosing a Spencer product

## 1 GENERAL INFORMATION





### 1.1 AIM AND CONTENTS

The aim of this manual is to supply all the information necessary so that the client, will not only attain adequate use of the appliance, he will also be capable of using the instrument in the most autonomous and secure way possible. This includes information regarding technical aspects, functioning, maintenance, spare parts and safety.

### 1.2 CONSERVATION OF THE INSTRUCTION MANUAL

The instruction and maintenance manual must be kept with the product, inside the specially provided container and above all, away from any substances or liquids which could compromise perfect legibility.

### 1.3 SYMBOLS USED

SYMBOL	MEANING
	General or specific warning
	See instructions for use
	Lot number
	Product code
SN	Serial number

### 1.4 SERVICING REQUESTS

For any information regarding the use, maintenance and installation, please contact the Spencer Customer Care Service on 0039 0521 541111 - fax 0039 0521 541222 - e-mail: info@spencer.it or write to Spencer Italia S.r.l. - Strada Cavi, 7 - 43044 Collecchio (Parma) - ITALY.

### 1.5 DEMOLITION

Follow the current regulations.

### 1.6 LABELLING

The serial number as indicated below can be found on each appliance and must not be removed or covered. In order to facilitate assistance please indicate or communicate the lot number (LOT) on the label.

## 2 WARNINGS

### 2.1 GENERAL WARNINGS

- Before carrying out any kind of operation on the appliance, the operator must carefully read the enclosed instructions, paying particular attention to the correct safety precautions and to the procedures to be followed for installation and for correct use.
- In the case of any doubts as to the correct interpretation of the instructions, please contact Spencer Italia S.r.l., for any necessary clarifications.
- Regularly check the appliance. Carry out the prescribed maintenance in order to keep the appliance in good condition and to guarantee correct functioning and a long life.
- In the case of any abnormalities or damage to the appliance, which could jeopardize the functioning and the safety, the appliance must be immediately removed from service.
- Do not alter or modify in any way the appliance; any such interference could cause malfunctions and injury to the patient and/or rescuer.
- The appliance must not in any way be tampered with. In such cases all responsibility will be denied for any malfunctions or injuries caused by the appliance itself. Those who modify or have modified, prepare or have prepared medical appliances in such a way that they no longer serve the purpose for which they were intended, or no longer supply the intended service, must satisfy the valid conditions for the introduction onto the market.
- Ensure that all the necessary precautions are taken in order to avoid the hazards that can arise as the result of contact with blood or body fluids.
- Handle with care.

### 2.2 SPECIFIC WARNINGS



- The product must be used by trained personnel only.
- If any failure or incorrect functioning of the device is detected, it must be immediately substituted with a similar item so that the rescue procedures are guaranteed without any interruption.
- When the device is being used, the assistance of qualified staff must be guaranteed.
- Always respect the maximum load capacity of the device, as indicated in this manual.
- Weight must be correctly distributed on the device as indicated in the present manual.
- Additional care must be taken when obstacles are present on the surface or the surface is not flat.

### 2.3 CONTRAINDICATIONS AND SIDE EFFECTS

The use of this device, if used as described in this manual, does not present any contra-indications or collateral effects.

### 2.4 PHYSICAL REQUIREMENTS OF THE OPERATORS



This device must be used by at least 4 sensible, strong operators with a good sense of balance during the loading/unloading phase. The capacities of the various operators must be considered before determining their roles in the use of the device.

### 3 DESCRIPTION OF PRODUCT


#### 3.1 INTENDED USE

WING is a trestle with variable heights in aluminium which is a useful aid in the chapel of rest of rest and offers an easy resting place for the coffin.

#### 3.2 MAIN COMPONENTS

1. Large aluminium resting surface
2. Narrow aluminium resting surface
3. Rubber prop
4. Alloy couplings

Fig. A



#### 3.3 MODELS

- ZC00405B Wing Aluminium trestle with adjustable heights, silver, 1 pc.  
ZC00406B Wing Aluminium trestle with adjustable heights, bronze, 1 pc.  
ZC00407B Wing Aluminium trestle with adjustable heights, silver, 2 pcs  
ZC00408B Wing Aluminium trestle with adjustable heights, bronze, 2 pcs

#### 3.4 TECHNICAL DATA

Minimum height	480 mm $\pm 3\%$
Maximum height	520 mm $\pm 3\%$
Minimum length	480 mm $\pm 3\%$
Maximum length	645 mm $\pm 3\%$
Width	490 mm
Weight for 2 pcs	6,5 kg
Load capacity for 2 pcs	270 kg
Structural material	Aluminium
Finishing	Polished / anodized

#### 3.5 ENVIRONMENTAL CONDITIONS

Functioning temperature: from -20 to +60 °C  
Storage temperature: from -20 to +60 °C  
Relative humidity: from 5 to 85%

### 4 OPERATING INSTRUCTIONS

#### 4.1 TRANSPORT AND STORAGE

Before transporting the appliance, make sure that it is correctly packaged ensuring also that there are no risks of shocks, bumps or falls during the transport itself. Keep the original packaging for use in case of any further transport. Damage to the appliance caused during transport and handling is not covered by the guarantee. Repairs or replacement of the damaged parts are the responsibility of the client. The appliance must be stored in a dry place free from humidity. During storage, do not place any heavy objects on top of the device; it should not be used as a support for anything apart for its specific function.

#### 4.2 PREPARATION

On receipt of the product:

- Remove the packaging and display the material so that all components are visible.
- Check that all the components/pieces on the accompanying list are present.

The appliance must be checked before every use so as to reveal any working abnormalities and/or damage caused by transport and/or storage.

#### 4.3 FUNCTIONING

Before use carefully check the following:

- General functioning of the device
- Tightening of screws and bolts

#### 4.3.1 LOADING THE COFFIN FOR USE OF THE DEVICE AS A WORK BENCH



**This phase requires at least four operators in good physical condition.**

1. Position both trestles on a horizontal flat ground surface.
2. Position the trestles at the desired height making sure that the narrow rests are facing each other. The distance between the two sides must be at least 150 cm.
3. Load the coffin on the top surface of the device making sure that the weight is evenly distributed.

#### 4.3.2 UNLOADING THE COFFIN



**This phase requires at least four operators in good physical condition.**

Lift the coffin and re positioned as desired.

#### 4.3.3 HEIGHT ADJUSTMENT



**This phase must be carried out without any weight being loaded on the device.**

Three different positions can be obtained:

- High (fig. B)
- Sloping (fig. C)
- Low (fig. D)



**Before resting the coffin, check to make sure the device is opened up correctly.**



Fig. B



Fig. C



Fig. D

4.4 TROUBLESHOOTING

PROBLEM	CAUSE	REMEDY
The turning system of the tressel is blocked	Excess weight loaded, wearing or insufficient lubrication	Clean and lubricate where necessary with the correct products; if the problem is not resolved contact the assistance centre
The rubber parts do not absorb the load	Excessive load or wearing	Immediately remove the device from use and contact the service centre

5 MAINTENANCE AND CLEANING

5.1 CLEANING

Wash the device with a clean cloth using mild detergents for sanitary purpose. Do not use any harsh detergent or solvent of any kind. If necessary dry using a soft cloth or chammy. The use of high pressure water systems will increase the probability of corrosion.

5.2 MAINTENANCE

5.2.1 Precautionary Maintenance

The person who carries out the precautionary maintenance of the appliance (user in person, manufacturer/supplier or a third party) has to guarantee the following basic requirements:  
Technical knowledge of the appliance and of the periodic maintenance procedures as described in these instructions.  
Specific qualifications and training in the maintenance operations of the appliance in question.  
The use of components/replacement parts/accessories that are either original or approved by the supplier, in such a way that each operation causes no alteration or modification to the appliance.  
Possession of the checklist of operations carried out on the appliance.  
We advice that a routine check programme is completed for the following:  
Lubrication of moving parts  
Tightening of screws and bolts  
Routine cleaning  
The device must be serviced by the manufacturer or by an authorised centre once a year.  
For any operations that are not carried out directly by the manufacturer but by an authorised centre, we have to underline that a report regarding all operations carried out must be requested. This will permit both Spencer Italia S.r.l. and the end user to keep a log book regarding the operations carried out on the device.

5.2.2 Special servicing



**Only the manufacturer or centres with written authorisation are authorised to complete any special servicing operations.**  
For any operations that are not carried out directly by the manufacturer but by an authorised centre, we have to underline that a report regarding all operations carried out must be requested. This will permit both Spencer Italia S.r.l. and the end user to keep a log book regarding the operations carried out on the device. The device, if used as indicated in the following instruction manual, has an average life span of 5 years.  
The life span can be expanded only following a general revision of the product that must be carried out by the manufacturer or by a centre authorised by the manufacturer.

6 ACCESSORIES AND SPARE PARTS

There are no accessories or replacement parts for this item.



Warning

The information contained in this document could be modified without any warning and is not to be intended as a commitment on behalf of Spencer Italia S.r.l. Spencer products are exported to many countries and the same identical regulations are not always valid. For this reason there could be differences between the description here described and the product actually delivered. Spencer continually strives to reach the perfection of all items sold. We therefore hope you will understand if we reserve the right, at any time, to modify the shape, equipment, lay-out or technical aspects that are herein described.